



JSC NTC "Metrotek"
52 Ectrozavodskaya str.
Moscow
Russia

Page 1 of 1

Certificate of Compliance

Electromagnetic Compatibility

For

Gigabit Ethernet Analyzer

Date of Issue: 2nd June 2010

SGS Reference: CST133541/1/AP/10

Details of Sample: Model No.: BERcut-ET/ETL
Unique Identifier: 8501

Date of Receipt: 1st February 2010

Date of Test: 7th February 2010

Test Required: Test for compliance with: -
EN55022:2006, A1:2007
EN55024:1998, A1:2001, A2:2003
EN61000-3-2:2006
EN61000-3-3:1995, A1:2001, A2:2005

Conclusion: In the configuration tested the sample meets the requirements of the standards detailed above.

Issue Number: 1

This document is issued by the Company under its General Conditions of Service accessible at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test certificate refer only to the sample(s) tested and such sample(s) are retained for 4 weeks only. This document cannot be reproduced except in full, without prior approval of the Company.



Authorized Signatory

A. Pers
Operations Manager



SGSPAPER
10063202

